

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN068973.D
 Acq On : 18 Oct 2021 12:38
 Operator : JC/MD
 Sample : M4243-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T3-2-TOP-GRAB

Quant Time: Oct 19 06:34:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	347417	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	598049	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	570258	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	205620	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	222951	48.673	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.340%
35) Dibromofluoromethane	8.024	113	176672	50.775	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.560%
50) Toluene-d8	10.443	98	755994	54.302	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.600%
62) 4-Bromofluorobenzene	12.733	95	262596	49.993	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.980%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	22541	13.187	ug/l	99
20) Methylene Chloride	5.103	84	11806	3.001	ug/l	98
43) Isopropyl Acetate	8.563	43	42242	4.538	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
Data File : VN068973.D
Acq On : 18 Oct 2021 12:38
Operator : JC/MD
Sample : M4243-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T3-2-TOP-GRAB

Quant Time: Oct 19 06:34:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

