

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN068975.D
 Acq On : 18 Oct 2021 13:28
 Operator : JC/MD
 Sample : M4243-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T3-4-TOP-GRAB

Quant Time: Oct 19 06:35:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	340083	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	579707	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	552305	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	199436	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.439	65	217857	48.587	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.180%
35) Dibromofluoromethane	8.024	113	170290	50.489	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.980%
50) Toluene-d8	10.443	98	731593	54.212	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.420%
62) 4-Bromofluorobenzene	12.733	95	254946	50.073	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.140%
Target Compounds						
					Qvalue	
16) Acetone	4.301	43	20376	12.177	ug/l	91
20) Methylene Chloride	5.100	84	12955	3.364	ug/l	91
43) Isopropyl Acetate	8.563	43	39661	4.395	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
Data File : VN068975.D
Acq On : 18 Oct 2021 13:28
Operator : JC/MD
Sample : M4243-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T3-4-TOP-GRAB

Quant Time: Oct 19 06:35:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

