

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN068977.D
 Acq On : 18 Oct 2021 14:18
 Operator : JC/MD
 Sample : M4243-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T3-2-BOTTOM-GRAB

Quant Time: Oct 19 06:35:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	348064	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	601392	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	584535	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	213384	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	226791	49.419	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.840%
35) Dibromofluoromethane	8.024	113	178995	51.157	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.320%
50) Toluene-d8	10.443	98	759680	54.263	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.520%
62) 4-Bromofluorobenzene	12.733	95	267516	50.647	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.300%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	22297	13.020	ug/l	97
20) Methylene Chloride	5.101	84	12019	3.049	ug/l	91
43) Isopropyl Acetate	8.563	43	48346	5.165	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
Data File : VN068977.D
Acq On : 18 Oct 2021 14:18
Operator : JC/MD
Sample : M4243-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T3-2-BOTTOM-GRAB

Quant Time: Oct 19 06:35:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

