

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN068979.D
 Acq On : 18 Oct 2021 15:09
 Operator : JC/MD
 Sample : M4243-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T3-4-BOTTOM-GRAB

Quant Time: Oct 19 06:36:38 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	348919	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	594789	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	568947	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	213079	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	223912	48.673	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.340%
35) Dibromofluoromethane	8.024	113	177156	51.193	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.380%
50) Toluene-d8	10.443	98	755443	54.560	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	109.120%
62) 4-Bromofluorobenzene	12.733	95	263187	50.380	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.760%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	21526	12.539	ug/l	100
20) Methylene Chloride	5.101	84	12260	3.103	ug/l	95
43) Isopropyl Acetate	8.563	43	53452	5.774	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
Data File : VN068979.D
Acq On : 18 Oct 2021 15:09
Operator : JC/MD
Sample : M4243-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T3-4-BOTTOM-GRAB

Quant Time: Oct 19 06:36:38 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

