

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN068981.D
 Acq On : 18 Oct 2021 15:59
 Operator : JC/MD
 Sample : M4243-13
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T4-2-TOP-GRAB

Quant Time: Oct 19 06:37:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	343648	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	591228	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	563137	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	205295	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	219986	48.553	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.100%
35) Dibromofluoromethane	8.024	113	176255	51.240	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.480%
50) Toluene-d8	10.443	98	747707	54.326	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.660%
62) 4-Bromofluorobenzene	12.734	95	254353	48.983	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.960%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	20496	12.122	ug/l	98
20) Methylene Chloride	5.103	84	12340	3.171	ug/l	95
43) Isopropyl Acetate	8.563	43	46950	5.102	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
Data File : VN068981.D
Acq On : 18 Oct 2021 15:59
Operator : JC/MD
Sample : M4243-13
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T4-2-TOP-GRAB

Quant Time: Oct 19 06:37:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

