

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN068982.D
 Acq On : 18 Oct 2021 16:24
 Operator : JC/MD
 Sample : M4243-14
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T4-3-TOP-GRAB

Quant Time: Oct 19 06:37:35 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	337123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	583804	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	565888	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	199746	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	215712	48.531	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.060%
35) Dibromofluoromethane	8.024	113	173779	51.162	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.320%
50) Toluene-d8	10.443	98	736858	54.219	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.440%
62) 4-Bromofluorobenzene	12.733	95	257205	50.162	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.320%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	19881	11.986	ug/l	98
20) Methylene Chloride	5.103	84	12479	3.269	ug/l	94
43) Isopropyl Acetate	8.566	43	54313	5.977	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
Data File : VN068982.D
Acq On : 18 Oct 2021 16:24
Operator : JC/MD
Sample : M4243-14
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T4-3-TOP-GRAB

Quant Time: Oct 19 06:37:35 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

