

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN068995.D
 Acq On : 18 Oct 2021 22:42
 Operator : JC/MD
 Sample : VN1018WBL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1018WBL02

Quant Time: Oct 19 09:57:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	328305	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	567741	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	524664	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	191760	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	211733	48.915	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.840%
35) Dibromofluoromethane	8.024	113	168885	51.128	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.260%
50) Toluene-d8	10.443	98	698734	52.868	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.740%
62) 4-Bromofluorobenzene	12.734	95	245667	49.267	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
Data File : VN068995.D
Acq On : 18 Oct 2021 22:42
Operator : JC/MD
Sample : VN1018WBL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1018WBL02

Quant Time: Oct 19 09:57:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

