

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN068999.D
 Acq On : 19 Oct 2021 00:23
 Operator : JC/MD
 Sample : M4268-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TRIP-BLANK

Manual Integrations
 APPROVED

MMDadoda
 10/19/2021 5:23:40 PM

Quant Time: Oct 19 09:58:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	327044	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	563222	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	513621	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	183499	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	207439	48.108	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.220%		
35) Dibromofluoromethane	8.024	113	165778	50.590	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	101.180%		
50) Toluene-d8	10.443	98	693951	52.928	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	105.860%		
62) 4-Bromofluorobenzene	12.733	95	236037	47.716	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	95.440%		
Target Compounds					Qvalue	
16) Acetone	4.304	43	3775m	2.346	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
Data File : VN068999.D
Acq On : 19 Oct 2021 00:23
Operator : JC/MD
Sample : M4268-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TRIP-BLANK

Manual Integrations
APPROVED

MMDadoda
10/19/2021 5:23:40 PM

Quant Time: Oct 19 09:58:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

