

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN069000.D
 Acq On : 19 Oct 2021 00:48
 Operator : JC/MD
 Sample : M4268-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 P257-DITCH

Quant Time: Oct 19 09:58:48 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	336019	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	582252	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	527668	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	188909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	212937	48.064	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.120%
35) Dibromofluoromethane	8.024	113	168460	49.728	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.460%
50) Toluene-d8	10.443	98	710770	52.439	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.880%
62) 4-Bromofluorobenzene	12.733	95	243288	47.574	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.140%
Target Compounds						
16) Acetone	4.293	43	6504	3.934	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
Data File : VN069000.D
Acq On : 19 Oct 2021 00:48
Operator : JC/MD
Sample : M4268-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
P257-DITCH

Quant Time: Oct 19 09:58:48 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

