

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN069001.D
 Acq On : 19 Oct 2021 1:13
 Operator : JC/MD
 Sample : M4268-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SW-RR-03-(SA)

Quant Time: Oct 19 09:58:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	330196	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	565869	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	504548	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	188685	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	209911	48.216	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.440%
35) Dibromofluoromethane	8.024	113	168123	51.066	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.140%
50) Toluene-d8	10.443	98	692724	52.587	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.180%
62) 4-Bromofluorobenzene	12.733	95	232448	46.770	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	93.540%
Target Compounds						
					Qvalue	
16) Acetone	4.304	43	6260	3.853	ug/l	93
52) Toluene	10.510	92	3947	0.417	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
Data File : VN069001.D
Acq On : 19 Oct 2021 1:13
Operator : JC/MD
Sample : M4268-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SW-RR-03-(SA)

Quant Time: Oct 19 09:58:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

