

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN069007.D
 Acq On : 19 Oct 2021 3:44
 Operator : JC/MD
 Sample : M4242-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T1-3-TOP-GRAB

Quant Time: Oct 19 09:59:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	339330	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	586220	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	560879	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	207090	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.437	65	221457	49.499	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.000%
35) Dibromofluoromethane	8.021	113	174336	51.115	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.220%
50) Toluene-d8	10.443	98	731987	53.638	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.280%
62) 4-Bromofluorobenzene	12.734	95	259807	50.460	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.920%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	20701	12.399	ug/l	99
18) Methyl Acetate	4.865	43	8065	1.274	ug/l	93
20) Methylene Chloride	5.098	84	11641	3.029	ug/l	96
43) Isopropyl Acetate	8.563	43	51871	5.685	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
Data File : VN069007.D
Acq On : 19 Oct 2021 3:44
Operator : JC/MD
Sample : M4242-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T1-3-TOP-GRAB

Quant Time: Oct 19 09:59:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

