

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN069013.D
 Acq On : 19 Oct 2021 6:15
 Operator : JC/MD
 Sample : M4242-12
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T2-1-TOP-GRAB

Quant Time: Oct 19 10:00:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	340818	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	586711	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	563580	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	207801	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	217359	48.371	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	96.740%
35) Dibromofluoromethane	8.024	113	174064	50.992	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.980%
50) Toluene-d8	10.443	98	740411	54.210	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.420%
62) 4-Bromofluorobenzene	12.733	95	260499	50.552	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.100%
Target Compounds						
					Qvalue	
16) Acetone	4.291	43	20807	12.408	ug/l	96
20) Methylene Chloride	5.101	84	10935	2.833	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
Data File : VN069013.D
Acq On : 19 Oct 2021 6:15
Operator : JC/MD
Sample : M4242-12
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T2-1-TOP-GRAB

Quant Time: Oct 19 10:00:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

