

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN068994.D
 Acq On : 18 Oct 2021 21:52
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 19 09:57:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.546	0.522	4.4	72	0.00
3 P	Chloromethane	0.611	0.605	1.0	78	0.00
4 C	Vinyl Chloride	0.621	0.711	-14.5#	87	0.00
5 T	Bromomethane	0.341	0.458	-34.3#	99	-0.02
6 T	Chloroethane	0.366	0.436	-19.1	91	-0.01
7 T	Trichlorofluoromethane	0.869	0.890	-2.4	80	0.00
8 T	Diethyl Ether	0.324	0.337	-4.0	82	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.467	0.495	-6.0	83	0.00
10 T	Methyl Iodide	0.678	0.653	3.7	75	0.00
11 T	Tert butyl alcohol	0.120	0.125	-4.2	86	0.00
12 CM	1,1-Dichloroethene	0.483	0.483	0.0	82	0.00
13 T	Acrolein	0.053	0.062	-17.0	106	0.00
14 T	Allyl chloride	0.866	0.846	2.3	75	0.00
15 T	Acrylonitrile	0.270	0.292	-8.1	83	0.00
16 T	Acetone	0.246	0.238	3.3	75	0.00
17 T	Carbon Disulfide	1.311	1.278	2.5	76	0.00
18 T	Methyl Acetate	0.933	1.048	-12.3	88	0.00
19 T	Methyl tert-butyl Ether	1.799	1.902	-5.7	81	0.00
20 T	Methylene Chloride	0.566	0.582	-2.8	83	0.00
21 T	trans-1,2-Dichloroethene	0.522	0.524	-0.4	80	0.00
22 T	Diisopropyl ether	1.801	1.895	-5.2	79	0.00
23 T	Vinyl Acetate	1.362	1.434	-5.3	79	0.00
24 P	1,1-Dichloroethane	0.979	1.019	-4.1	82	0.00
25 T	2-Butanone	0.372	0.402	-8.1	82	0.00
26 T	2,2-Dichloropropane	0.933	0.830	11.0	70	0.00
27 T	cis-1,2-Dichloroethene	0.631	0.637	-1.0	80	0.00
28 T	Bromochloromethane	0.443	0.507	-14.4	94	0.00
29 T	Tetrahydrofuran	0.241	0.271	-12.4	84	0.00
30 C	Chloroform	1.031	1.108	-7.5#	83	0.00
31 T	Cyclohexane	1.025	0.953	7.0	77	0.00
32 T	1,1,1-Trichloroethane	0.961	0.997	-3.7	81	0.00
33 S	1,2-Dichloroethane-d4	0.659	0.742	-12.6	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.291	0.318	-9.3	101	0.00
36 T	1,1-Dichloropropene	0.427	0.442	-3.5	82	0.00
37 T	Ethyl Acetate	0.441	0.468	-6.1	84	0.00
38 T	Carbon Tetrachloride	0.469	0.476	-1.5	81	0.00
39 T	Methylcyclohexane	0.550	0.547	0.5	78	0.00
40 TM	Benzene	1.300	1.333	-2.5	81	0.00
41 T	Methacrylonitrile	0.222	0.236	-6.3	82	0.00
42 TM	1,2-Dichloroethane	0.461	0.486	-5.4	83	0.00
43 T	Isopropyl Acetate	0.778	0.839	-7.8	84	0.00
44 TM	Trichloroethene	0.325	0.328	-0.9	80	0.00
45 C	1,2-Dichloropropane	0.337	0.349	-3.6#	82	0.00
46 T	Dibromomethane	0.230	0.242	-5.2	84	0.00
47 T	Bromodichloromethane	0.477	0.500	-4.8	82	0.00
48 T	Methyl methacrylate	0.366	0.352	3.8	75	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
 Data File : VN068994.D
 Acq On : 18 Oct 2021 21:52
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 19 09:57:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.007	-16.7	89	0.00
50 S	Toluene-d8	1.164	1.294	-11.2	101	0.00
51 T	4-Methyl-2-Pentanone	0.451	0.498	-10.4	85	0.00
52 CM	Toluene	0.837	0.891	-6.5#	82	0.00
53 T	t-1,3-Dichloropropene	0.541	0.555	-2.6	80	0.00
54 T	cis-1,3-Dichloropropene	0.563	0.573	-1.8	80	0.00
55 T	1,1,2-Trichloroethane	0.330	0.352	-6.7	85	0.00
56 T	Ethyl methacrylate	0.554	0.594	-7.2	84	0.00
57 T	1,3-Dichloropropane	0.544	0.590	-8.5	84	0.00
58 T	2-Chloroethyl Vinyl ether	0.134	0.172	-28.4#	101	0.00
59 T	2-Hexanone	0.324	0.365	-12.7	87	0.00
60 T	Dibromochloromethane	0.372	0.383	-3.0	81	0.00
61 T	1,2-Dibromoethane	0.338	0.366	-8.3	83	0.00
62 S	4-Bromofluorobenzene	0.439	0.499	-13.7	104	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	88	0.00
64 T	Tetrachloroethene	0.322	0.299	7.1	74	0.00
65 PM	Chlorobenzene	0.946	0.999	-5.6	85	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.381	-3.0	82	0.00
67 C	Ethyl Benzene	1.758	1.868	-6.3#	84	0.00
68 T	m/p-Xylenes	0.666	0.697	-4.7	83	0.00
69 T	o-Xylene	0.667	0.701	-5.1	83	0.00
70 T	Styrene	1.086	1.182	-8.8	85	0.00
71 P	Bromoform	0.297	0.296	0.3	79	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	4.162	4.262	-2.4	85	0.00
74 T	N-amyl acetate	1.577	1.716	-8.8	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.203	1.311	-9.0	90	0.00
76 T	1,2,3-Trichloropropane	1.064	1.171	-10.1	85	0.00
77 T	Bromobenzene	0.920	0.931	-1.2	83	0.00
78 T	n-propylbenzene	4.677	4.894	-4.6	85	0.00
79 T	2-Chlorotoluene	2.883	2.990	-3.7	86	0.00
80 T	1,3,5-Trimethylbenzene	3.466	3.552	-2.5	84	0.00
81 T	trans-1,4-Dichloro-2-butene	0.445	0.432	2.9	82	0.00
82 T	4-Chlorotoluene	2.786	2.930	-5.2	86	0.00
83 T	tert-Butylbenzene	3.018	3.066	-1.6	84	0.00
84 T	1,2,4-Trimethylbenzene	3.340	3.472	-4.0	85	0.00
85 T	sec-Butylbenzene	4.179	4.375	-4.7	85	0.00
86 T	p-Isopropyltoluene	3.352	3.518	-5.0	85	0.00
87 T	1,3-Dichlorobenzene	1.592	1.655	-4.0	84	0.00
88 T	1,4-Dichlorobenzene	1.580	1.627	-3.0	83	0.00
89 T	n-Butylbenzene	2.791	2.989	-7.1	86	0.00
90 T	Hexachloroethane	0.704	0.711	-1.0	83	0.00
91 T	1,2-Dichlorobenzene	1.546	1.608	-4.0	83	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.232	0.254	-9.5	87	0.00
93 T	1,2,4-Trichlorobenzene	0.686	0.753	-9.8	83	0.00
94 T	Hexachlorobutadiene	0.440	0.402	8.6	74	0.00
95 T	Naphthalene	1.793	2.152	-20.0	88	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101821\
Data File : VN068994.D
Acq On : 18 Oct 2021 21:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 19 09:57:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.650	0.728	-12.0	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5