

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101918\
 Data File : VN051944.D
 Acq On : 19 Oct 2018 13:33
 Operator : MD\SY
 Sample : PB113937ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 11 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 PB113937ZHE#05

Quant Time: Oct 19 23:58:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101718W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 19 05:11:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	967231	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1407996	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1215452	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	406166	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	522756	58.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.60%	
35) Dibromofluoromethane	7.59	113	523484	58.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.98%	
50) Toluene-d8	10.09	98	1911872	58.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	117.38%	
62) 4-Bromofluorobenzene	12.40	95	523815	48.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.36%	
Target Compounds						
16) Acetone	3.83	43	10436	6.573	ug/l #	88
20) Methylene Chloride	4.56	84	11418	1.415	ug/l	92
43) Isopropyl Acetate	8.17	43	300622	29.745	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN101918\
Data File : VN051944.D
Acq On : 19 Oct 2018 13:33
Operator : MD\SY
Sample : PB113937ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 11 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
PB113937ZHE#05

Quant Time: Oct 19 23:58:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101718W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 19 05:11:28 2018
Response via : Initial Calibration

