

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101918\
 Data File : VN051954.D
 Acq On : 19 Oct 2018 18:01
 Operator : MD\SY
 Sample : J5538-23 50X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 BS01

Manual Integrations
 APPROVED

MMDadoda
 10/22/2018 11:35:47 AM

Quant Time: Oct 20 00:30:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101718W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 19 05:11:28 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	962196	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	1414910	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	1312171	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	648295	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	540420	60.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.16%	
35) Dibromofluoromethane	7.59	113	1834	0.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	0.40%	
50) Toluene-d8	10.09	98	1956104	59.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	119.52%	
62) 4-Bromofluorobenzene	12.40	95	631465	57.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.58%	
Target Compounds						
16) Acetone	3.82	43	15823m	10.018	ug/l	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN101918\
Data File : VN051954.D
Acq On : 19 Oct 2018 18:01
Operator : MD\SY
Sample : J5538-23 50X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
BS01

Manual Integrations
APPROVED

MMDadoda
10/22/2018 11:35:47 AM

Quant Time: Oct 20 00:30:50 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101718W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 19 05:11:28 2018
Response via : Initial Calibration

