


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN101918\
Data File : VN051934.D
Acq On : 19 Oct 2018 7:34
Operator : MD\SY
Sample : BFB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 1 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101718W.M
Title : SW846 8260
Last Update : Fri Oct 19 05:11:28 2018



AutoFind: Scans 3304, 3305, 3306; Background Corrected with Scan 3290

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	13175	PASS
75	95	30	60	52.1	41634	PASS
95	95	100	100	100.0	79885	PASS
96	95	5	9	6.7	5338	PASS
173	174	0.00	2	0.4	277	PASS
174	95	50	100	99.0	79061	PASS
175	174	5	9	7.4	5824	PASS
176	174	95	101	95.2	75256	PASS
177	176	5	9	7.0	5242	PASS