

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
 Data File : VN069028.D
 Acq On : 19 Oct 2021 16:09
 Operator : JC/MD
 Sample : M4251-09RE
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TBRE

Quant Time: Oct 20 05:48:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	337990	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	579893	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	537243	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	191339	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	217795	48.874	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.740%
35) Dibromofluoromethane	8.024	113	172553	51.144	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	102.280%
50) Toluene-d8	10.443	98	717090	53.120	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.240%
62) 4-Bromofluorobenzene	12.733	95	242491	47.611	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.220%
Target Compounds						
16) Acetone	4.309	43	3172	1.907	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
Data File : VN069028.D
Acq On : 19 Oct 2021 16:09
Operator : JC/MD
Sample : M4251-09RE
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TBRE

Quant Time: Oct 20 05:48:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

