

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
 Data File : VN069030.D
 Acq On : 19 Oct 2021 16:59
 Operator : JC/MD
 Sample : PB139960TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB139960TB

Quant Time: Oct 20 05:49:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	336707	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.968	114	586895	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	565269	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	210728	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	219661	49.480	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.960%
35) Dibromofluoromethane	8.024	113	173031	50.674	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.340%
50) Toluene-d8	10.443	98	735421	53.828	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.660%
62) 4-Bromofluorobenzene	12.734	95	262790	50.981	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.960%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	16646	10.048	ug/l	97
20) Methylene Chloride	5.101	84	17541	4.600	ug/l	92
43) Isopropyl Acetate	8.563	43	35252	3.859	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
Data File : VN069030.D
Acq On : 19 Oct 2021 16:59
Operator : JC/MD
Sample : PB139960TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB139960TB

Quant Time: Oct 20 05:49:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

