

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
 Data File : VN069031.D
 Acq On : 19 Oct 2021 17:24
 Operator : JC/MD
 Sample : M4242-17
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T2-2-BOTTOM-GRAB

Quant Time: Oct 20 05:49:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	329547	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	576460	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	543673	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	203149	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	216136	49.744	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.480%
35) Dibromofluoromethane	8.024	113	170542	50.849	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.700%
50) Toluene-d8	10.443	98	715378	53.309	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.620%
62) 4-Bromofluorobenzene	12.734	95	252823	49.935	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.880%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	19352	11.935	ug/l	97
20) Methylene Chloride	5.103	84	12532	3.358	ug/l	99
43) Isopropyl Acetate	8.566	43	42213	4.705	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
Data File : VN069031.D
Acq On : 19 Oct 2021 17:24
Operator : JC/MD
Sample : M4242-17
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T2-2-BOTTOM-GRAB

Quant Time: Oct 20 05:49:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

