

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
 Data File : VN069032.D
 Acq On : 19 Oct 2021 17:49
 Operator : JC/MD
 Sample : M4242-18
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T2-3-BOTTOM-GRAB

Quant Time: Oct 20 05:49:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	328006	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	567033	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	548007	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	196483	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	210311	48.631	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.260%
35) Dibromofluoromethane	8.027	113	166568	50.490	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.980%
50) Toluene-d8	10.443	98	711473	53.899	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.800%
62) 4-Bromofluorobenzene	12.733	95	249442	50.087	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.180%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	22792	14.123	ug/l	97
20) Methylene Chloride	5.101	84	10955	2.949	ug/l	89
43) Isopropyl Acetate	8.563	43	47010	5.326	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
Data File : VN069032.D
Acq On : 19 Oct 2021 17:49
Operator : JC/MD
Sample : M4242-18
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T2-3-BOTTOM-GRAB

Quant Time: Oct 20 05:49:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

