

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
 Data File : VN069034.D
 Acq On : 19 Oct 2021 18:39
 Operator : JC/MD
 Sample : M4244-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 T5-1-BOTTOM-GRAB

Quant Time: Oct 20 05:49:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	340156	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	592850	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	550522	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	206259	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	221157	49.312	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.620%
35) Dibromofluoromethane	8.024	113	174290	50.530	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.060%
50) Toluene-d8	10.443	98	734396	53.213	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.420%
62) 4-Bromofluorobenzene	12.733	95	250675	48.142	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.280%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	28867	17.248	ug/l	98
18) Methyl Acetate	4.873	43	8399	1.323	ug/l #	90
20) Methylene Chloride	5.103	84	10314	2.677	ug/l	94
43) Isopropyl Acetate	8.566	43	45889	4.973	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
Data File : VN069034.D
Acq On : 19 Oct 2021 18:39
Operator : JC/MD
Sample : M4244-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
T5-1-BOTTOM-GRAB

Quant Time: Oct 20 05:49:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

