

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
 Data File : VN069038.D
 Acq On : 19 Oct 2021 20:20
 Operator : JC/MD
 Sample : M4222-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TEST-A(1)

Quant Time: Oct 20 05:50:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	333249	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	579115	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	538229	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	202287	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	218185	49.658	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.320%
35) Dibromofluoromethane	8.024	113	171546	50.914	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.820%
50) Toluene-d8	10.443	98	717922	53.253	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.500%
62) 4-Bromofluorobenzene	12.733	95	242579	47.692	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	95.380%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	21559	13.148	ug/l	99
18) Methyl Acetate	4.870	43	7820	1.257	ug/l	91
20) Methylene Chloride	5.103	84	13012	3.448	ug/l	98
43) Isopropyl Acetate	8.563	43	43251	4.798	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
Data File : VN069038.D
Acq On : 19 Oct 2021 20:20
Operator : JC/MD
Sample : M4222-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TEST-A(1)

Quant Time: Oct 20 05:50:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

