

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
 Data File : VN069039.D
 Acq On : 19 Oct 2021 20:45
 Operator : JC/MD
 Sample : M4224-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 515-Z

Manual Integrations
 APPROVED

MMDadoda
 10/20/2021 5:47:25 PM

Quant Time: Oct 20 05:50:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	328338	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	570863	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	543059	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	203323	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	216620	50.039	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.080%			
35) Dibromofluoromethane	8.024	113	169130	50.922	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 101.840%			
50) Toluene-d8	10.443	98	716261	53.898	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 107.800%			
62) 4-Bromofluorobenzene	12.733	95	248197	49.502	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 99.000%			
Target Compounds					Qvalue	
16) Acetone	4.293	43	20403	12.629	ug/l	100
18) Methyl Acetate	4.870	43	8579m	1.400	ug/l	
20) Methylene Chloride	5.106	84	13112	3.526	ug/l	98
43) Isopropyl Acetate	8.566	43	50297	5.661	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
Data File : VN069039.D
Acq On : 19 Oct 2021 20:45
Operator : JC/MD
Sample : M4224-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
515-Z

Manual Integrations
APPROVED

MMDadoda
10/20/2021 5:47:25 PM

Quant Time: Oct 20 05:50:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

