

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
 Data File : VN069044.D
 Acq On : 20 Oct 2021 00:06
 Operator : JC/MD
 Sample : VN1019WBL02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1019WBL02

Quant Time: Oct 20 05:51:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	331120	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	581280	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	528378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	190964	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	215622	49.390	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.780%
35) Dibromofluoromethane	8.024	113	170337	50.367	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.740%
50) Toluene-d8	10.443	98	714202	52.780	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.560%
62) 4-Bromofluorobenzene	12.734	95	241049	47.215	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	94.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
Data File : VN069044.D
Acq On : 20 Oct 2021 00:06
Operator : JC/MD
Sample : VN1019WBL02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1019WBL02

Quant Time: Oct 20 05:51:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

