



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101921\
 Data File : VN069042.D
 Acq On : 19 Oct 2021 22:25
 Operator : JC/MD
 Sample : BFB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Title : SW846 8260
 Last Update : Wed Oct 13 05:44:40 2021



AutoFind: Scans 4005, 4006, 4007; Background Corrected with Scan 3989

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.6	12656	PASS
75	95	30	60	52.2	33659	PASS
95	95	100	100	100.0	64525	PASS
96	95	5	9	6.8	4370	PASS
173	174	0.00	2	0.4	185	PASS
174	95	50	100	70.8	45685	PASS
175	174	5	9	7.5	3444	PASS
176	174	95	101	97.2	44384	PASS
177	176	5	9	6.7	2975	PASS