

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101922\
 Data File : VN074936.D
 Acq On : 19 Oct 2022 17:44
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Oct 20 06:19:48 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101922W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 20 06:14:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	426654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	787786	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	673433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	224226	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	296535	29.283	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	58.560%#		
35) Dibromofluoromethane	8.177	113	243062	28.131	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	56.260%#		
50) Toluene-d8	10.571	98	949567	27.093	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	54.180%#		
62) 4-Bromofluorobenzene	12.853	95	230071	24.029	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	48.060%#		
Target Compounds						Qvalue
73) Isopropylbenzene	12.688	105	4742	0.231	ug/l	92
74) N-amyl acetate	12.494	43	5008	0.478	ug/l #	90
77) Bromobenzene	12.976	156	1946	0.426	ug/l	72
78) n-propylbenzene	13.035	91	8015	0.347	ug/l	89
79) 2-Chlorotoluene	13.129	91	5931	0.458	ug/l	89
80) 1,3,5-Trimethylbenzene	13.170	105	4843	0.305	ug/l	80
82) 4-Chlorotoluene	13.129	91	5931	0.458	ug/l	89
83) tert-Butylbenzene	13.447	119	3498	0.236	ug/l	95
84) 1,2,4-Trimethylbenzene	13.476	105	4621	0.305	ug/l	91
85) sec-Butylbenzene	13.618	105	5417	0.269	ug/l	95
86) p-Isopropyltoluene	13.729	119	5315	0.351	ug/l #	66
87) 1,3-Dichlorobenzene	13.741	146	6099	0.815	ug/l	84
88) 1,4-Dichlorobenzene	13.806	146	7701	1.032	ug/l #	30
89) n-Butylbenzene	14.053	91	7405	0.594	ug/l	95
91) 1,2-Dichlorobenzene	14.106	146	4545	0.610	ug/l	79
93) 1,2,4-Trichlorobenzene	15.394	180	7502	2.021	ug/l	93
95) Naphthalene	15.647	128	45611	3.286	ug/l	97
96) 1,2,3-Trichlorobenzene	15.841	180	10046	2.619	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101922\
Data File : VN074936.D
Acq On : 19 Oct 2022 17:44
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Oct 20 06:19:48 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101922W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 20 06:14:51 2022
Response via : Initial Calibration

