

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101923\
 Data File : VN079609.D
 Acq On : 19 Oct 2023 10:28
 Operator : JC\MD
 Sample : VN1019WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1019WBL02

Quant Time: Oct 20 01:26:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	255239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	492321	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	490870	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	160787	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	219430	53.022	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.040%
35) Dibromofluoromethane	8.171	113	192357	54.369	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.740%
50) Toluene-d8	10.570	98	595096	46.486	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.980%
62) 4-Bromofluorobenzene	12.853	95	203220	49.466	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101923\
Data File : VN079609.D
Acq On : 19 Oct 2023 10:28
Operator : JC\MD
Sample : VN1019WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1019WBL02

Quant Time: Oct 20 01:26:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

