

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101923\
 Data File : VN079613.D
 Acq On : 19 Oct 2023 12:50
 Operator : JC\MD
 Sample : 04919-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LSS-B2-COMP-101723

Quant Time: Oct 20 01:28:30 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	247108	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	494528	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	503260	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	152241	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	209386	52.259	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 104.520%	
35) Dibromofluoromethane	8.171	113	188639	53.080	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.160%	
50) Toluene-d8	10.571	98	616014	47.906	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.820%	
62) 4-Bromofluorobenzene	12.853	95	203804	49.387	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 98.780%	
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	19721	12.410	ug/l	92
20) Methylene Chloride	5.283	84	11118	2.624	ug/l #	88
43) Isopropyl Acetate	8.694	43	161501	17.369	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101923\
Data File : VN079613.D
Acq On : 19 Oct 2023 12:50
Operator : JC\MD
Sample : 04919-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LSS-B2-COMP-101723

Quant Time: Oct 20 01:28:30 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

