

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101923\
 Data File : VN079614.D
 Acq On : 19 Oct 2023 13:14
 Operator : JC\MD
 Sample : 04919-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LSS-B5-COMP-101723

Quant Time: Oct 20 01:28:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	230628	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	473234	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	493041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	155524	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	205167	54.866	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.740%	
35) Dibromofluoromethane	8.171	113	180531	53.084	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 106.160%	
50) Toluene-d8	10.570	98	590419	47.981	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.960%	
62) 4-Bromofluorobenzene	12.853	95	202711	51.332	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 102.660%	
Target Compounds						
						Qvalue
16) Acetone	4.441	43	16383	11.046	ug/l #	85
20) Methylene Chloride	5.277	84	10475	2.655	ug/l #	84
43) Isopropyl Acetate	8.694	43	165029	18.547	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101923\
Data File : VN079614.D
Acq On : 19 Oct 2023 13:14
Operator : JC\MD
Sample : 04919-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LSS-B5-COMP-101723

Quant Time: Oct 20 01:28:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

