

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101923\
 Data File : VN079615.D
 Acq On : 19 Oct 2023 13:38
 Operator : JC\MD
 Sample : 04920-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 LSS-B1-COMP-101623

Quant Time: Oct 20 01:29:17 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	239009	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	493635	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	507736	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	158764	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	213373	55.059	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.120%	
35) Dibromofluoromethane	8.171	113	185739	52.359	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 104.720%	
50) Toluene-d8	10.571	98	611036	47.605	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.200%	
62) 4-Bromofluorobenzene	12.853	95	209233	50.794	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.580%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	18503	12.038	ug/l	100
20) Methylene Chloride	5.289	84	11331	2.798	ug/l #	86
43) Isopropyl Acetate	8.694	43	184003	19.825	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101923\
Data File : VN079615.D
Acq On : 19 Oct 2023 13:38
Operator : JC\MD
Sample : 04920-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
LSS-B1-COMP-101623

Quant Time: Oct 20 01:29:17 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration

