

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101923\
 Data File : VN079609.D
 Acq On : 19 Oct 2023 10:28
 Operator : JC\MD
 Sample : VN1019WBL02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1019WBL02

Quant Time: Oct 20 01:26:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 06 03:34:49 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	255239	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	492321	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	490870	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	160787	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	219430	53.022	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	106.040%
35) Dibromofluoromethane	8.171	113	192357	54.369	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	108.740%
50) Toluene-d8	10.570	98	595096	46.486	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	92.980%
62) 4-Bromofluorobenzene	12.853	95	203220	49.466	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.940%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN101923\
Data File : VN079609.D
Acq On : 19 Oct 2023 10:28
Operator : JC\MD
Sample : VN1019WBL02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1019WBL02

Quant Time: Oct 20 01:26:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N100523W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 06 03:34:49 2023
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.229 min Scan# 1067
Delta R.T. -0.001 min
Lab File: VN079609.D
Acq: 19 Oct 2023 10:28




Instrument :
MSVOA_N
ClientSampleId :
VN1019WBL02


Tgt Ion:168 Resp: 255239
Ion Ratio Lower Upper
168 100
99 60.8 53.0 79.6



#33
1,2-Dichloroethane-d4
Concen: 53.022 ug/l
RT: 8.582 min Scan# 1127
Delta R.T. -0.000 min
Lab File: VN079609.D
Acq: 19 Oct 2023 10:28

Tgt Ion: 65 Resp: 219430
Ion Ratio Lower Upper
65 100
67 50.7 0.0 106.6









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.106 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN079609.D
Acq: 19 Oct 2023 10:28




Instrument :
MSVOA_N
ClientSampleId :
VN1019WBL02


Tgt Ion:114 Resp: 492321
Ion Ratio Lower Upper
114 100
63 20.8 0.0 36.4
88 16.0 0.0 30.6



#35
Dibromofluoromethane
Concen: 54.369 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. -0.000 min
Lab File: VN079609.D
Acq: 19 Oct 2023 10:28

Tgt Ion:113 Resp: 192357
Ion Ratio Lower Upper
113 100
111 104.8 82.6 123.8
192 17.3 12.7 19.1









#50
Toluene-d8
Concen: 46.486 ug/l
RT: 10.570 min Scan# 1465
Delta R.T. -0.000 min
Lab File: VN079609.D
Acq: 19 Oct 2023 10:28




Instrument :
MSVOA_N
ClientSampleId :
VN1019WBL02


Tgt Ion: 98 Resp: 595096
Ion Ratio Lower Upper
98 100
100 63.9 52.0 78.0



#62
4-Bromofluorobenzene
Concen: 49.466 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. -0.000 min
Lab File: VN079609.D
Acq: 19 Oct 2023 10:28

Tgt Ion: 95 Resp: 203220
Ion Ratio Lower Upper
95 100
174 76.0 0.0 143.8
176 73.1 0.0 135.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.870 min Scan# 117
Delta R.T. -0.000 min
Lab File: VN079609.D
Acq: 19 Oct 2023 10:28

Instrument :
MSVOA_N
ClientSampleId :
VN1019WBL02

Tgt Ion:117 Resp: 490870
Ion Ratio Lower Upper
117 100
82 58.6 44.0 66.0
119 31.9 25.6 38.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. -0.000 min
Lab File: VN079609.D
Acq: 19 Oct 2023 10:28

Tgt Ion:152 Resp: 160787
Ion Ratio Lower Upper
152 100
115 59.7 32.3 96.8
150 155.5 0.0 355.8

