

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102020\
 Data File : VN064196.D
 Acq On : 20 Oct 2020 15:25
 Operator : JC/MD
 Sample : VN1020WBL01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1020WBL01

Quant Time: Oct 21 04:58:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N102020W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Oct 20 11:51:57 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	39818	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	228185	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	197943	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.98	65	103810	31.84	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	106.13%
60) 4-Bromofluorobenzene	12.37	95	99792	28.13	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	93.77%
63) Toluene-d8	10.06	98	266390	29.73	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.10%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102020\
Data File : VN064196.D
Acq On : 20 Oct 2020 15:25
Operator : JC/MD
Sample : VN1020WBL01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1020WBL01

Quant Time: Oct 21 04:58:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N102020W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Oct 20 11:51:57 2020
Response via : Initial Calibration

