

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102020\
 Data File : VN064203.D
 Acq On : 20 Oct 2020 18:17
 Operator : JC/MD
 Sample : L4473-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-020-IDW/GW-20201019

Manual Integrations
 APPROVED

MMDadoda
 10/23/2020 2:28:57 PM

Quant Time: Oct 21 05:13:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N102020W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Tue Oct 20 11:51:57 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.15	128	48204	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.55	114	279908	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.38	117	245508	30.00	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.99	65	133933	33.94	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery =	113.13%		
60) 4-Bromofluorobenzene	12.38	95	123824	28.14	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery =	93.80%		
63) Toluene-d8	10.06	98	336472	30.27	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery =	100.90%		
Target Compounds						
22) cis-1,2-Dichloroethene	6.78	96	10124m	2.919	ug/l	
37) Trichloroethene	8.80	130	49211	13.808	ug/l	93
61) Tetrachloroethene	10.60	164	2711	0.819	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102020\
Data File : VN064203.D
Acq On : 20 Oct 2020 18:17
Operator : JC/MD
Sample : L4473-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-020-IDW/GW-20201019

Manual Integrations
APPROVED

MMDadoda
10/23/2020 2:28:57 PM

Quant Time: Oct 21 05:13:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N102020W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Tue Oct 20 11:51:57 2020
Response via : Initial Calibration

