

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
 Data File : VN069054.D
 Acq On : 20 Oct 2021 11:46
 Operator : JC/MD
 Sample : VN1020WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1020WBL01

Quant Time: Oct 20 17:12:06 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	334609	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	582063	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	536549	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	197207	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	218068	49.429	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.860%
35) Dibromofluoromethane	8.024	113	171281	50.578	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.160%
50) Toluene-d8	10.443	98	725102	53.513	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.020%
62) 4-Bromofluorobenzene	12.733	95	251883	49.271	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
Data File : VN069054.D
Acq On : 20 Oct 2021 11:46
Operator : JC/MD
Sample : VN1020WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1020WBL01

Quant Time: Oct 20 17:12:06 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

