

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
 Data File : VN069058.D
 Acq On : 20 Oct 2021 13:40
 Operator : JC/MD
 Sample : M4286-03
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SW-RR-03-(SA)

Quant Time: Oct 20 16:41:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	340601	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	597304	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	556320	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	199455	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	220016	48.994	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.980%
35) Dibromofluoromethane	8.024	113	174811	50.303	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.600%
50) Toluene-d8	10.443	98	735977	52.930	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.860%
62) 4-Bromofluorobenzene	12.733	95	255444	48.692	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.380%
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	5977	3.567	ug/l	98
52) Toluene	10.507	92	5796	0.580	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
Data File : VN069058.D
Acq On : 20 Oct 2021 13:40
Operator : JC/MD
Sample : M4286-03
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SW-RR-03-(SA)

Quant Time: Oct 20 16:41:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

