

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
 Data File : VN069061.D
 Acq On : 20 Oct 2021 14:56
 Operator : JC/MD
 Sample : M4259-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1-BD

Quant Time: Oct 20 23:14:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	326054	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	573021	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	555751	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	206042	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	210834	49.044	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.080%
35) Dibromofluoromethane	8.027	113	166356	49.898	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	99.800%
50) Toluene-d8	10.446	98	718053	53.829	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.660%
62) 4-Bromofluorobenzene	12.733	95	252221	50.115	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.240%
Target Compounds						
					Qvalue	
16) Acetone	4.301	43	13341	8.316	ug/l	99
18) Methyl Acetate	4.867	43	4339	0.713	ug/l #	66
20) Methylene Chloride	5.106	84	19618	5.313	ug/l	93
43) Isopropyl Acetate	8.566	43	42511	4.766	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
Data File : VN069061.D
Acq On : 20 Oct 2021 14:56
Operator : JC/MD
Sample : M4259-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1-BD

Quant Time: Oct 20 23:14:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

