

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
 Data File : VN069062.D
 Acq On : 20 Oct 2021 15:21
 Operator : JC/MD
 Sample : M4259-10
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1-AS

Quant Time: Oct 20 23:14:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	347748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	609464	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	586361	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	216914	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	226012	49.294	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	98.580%		
35) Dibromofluoromethane	8.024	113	177104	49.946	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	99.900%		
50) Toluene-d8	10.443	98	763735	53.830	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	107.660%		
62) 4-Bromofluorobenzene	12.733	95	270737	50.578	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	101.160%		
Target Compounds						Qvalue
16) Acetone	4.296	43	11782	6.886	ug/l	92
18) Methyl Acetate	4.865	43	6849	1.055	ug/l #	75
20) Methylene Chloride	5.109	84	18915	4.803	ug/l	92
43) Isopropyl Acetate	8.566	43	47227	4.978	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
Data File : VN069062.D
Acq On : 20 Oct 2021 15:21
Operator : JC/MD
Sample : M4259-10
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1-AS

Quant Time: Oct 20 23:14:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

