

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
 Data File : VN069063.D
 Acq On : 20 Oct 2021 15:46
 Operator : JC/MD
 Sample : M4249-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP1-(S)

Quant Time: Oct 20 23:14:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	344796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	603473	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	582260	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	209267	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	223847	49.240	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.480%
35) Dibromofluoromethane	8.024	113	176469	50.261	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.520%
50) Toluene-d8	10.446	98	756053	53.818	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.640%
62) 4-Bromofluorobenzene	12.734	95	265507	50.093	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.180%
Target Compounds						
					Qvalue	
16) Acetone	4.301	43	12048	7.102	ug/l	99
18) Methyl Acetate	4.870	43	8496	1.320	ug/l #	69
20) Methylene Chloride	5.098	84	15891	4.070	ug/l	95
43) Isopropyl Acetate	8.569	43	45325	4.825	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
Data File : VN069063.D
Acq On : 20 Oct 2021 15:46
Operator : JC/MD
Sample : M4249-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP1-(S)

Quant Time: Oct 20 23:14:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

