

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
 Data File : VN069064.D
 Acq On : 20 Oct 2021 16:11
 Operator : JC/MD
 Sample : M4249-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP1-(D)

Manual Integrations
 APPROVED

MMDadoda
 10/21/2021 6:55:23 PM

Quant Time: Oct 20 23:14:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	340992	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	597487	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.749	117	583749	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	214679	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	225871	50.240	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.480%			
35) Dibromofluoromethane	8.024	113	175480	50.480	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 100.960%			
50) Toluene-d8	10.443	98	757248	54.443	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 108.880%			
62) 4-Bromofluorobenzene	12.733	95	268955	51.252	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 102.500%			
Target Compounds					Qvalue	
16) Acetone	4.301	43	11273	6.719	ug/l	98
18) Methyl Acetate	4.875	43	8011m	1.259	ug/l	
20) Methylene Chloride	5.109	84	15405	3.989	ug/l	94
43) Isopropyl Acetate	8.568	43	64345	6.919	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
Data File : VN069064.D
Acq On : 20 Oct 2021 16:11
Operator : JC/MD
Sample : M4249-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP1-(D)

Manual Integrations
APPROVED

MMDadoda
10/21/2021 6:55:23 PM

Quant Time: Oct 20 23:14:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

