

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
 Data File : VN069067.D
 Acq On : 20 Oct 2021 17:27
 Operator : JC/MD
 Sample : M4252-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TEST-BA-(1)

Quant Time: Oct 20 23:15:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	335184	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	587097	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	560532	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	210525	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	219399	49.646	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.300%
35) Dibromofluoromethane	8.027	113	172804	50.590	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.180%
50) Toluene-d8	10.443	98	731474	53.521	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.040%
62) 4-Bromofluorobenzene	12.736	95	252455	48.959	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.920%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	11685	7.085	ug/l	97
18) Methyl Acetate	4.870	43	10623	1.698	ug/l	95
20) Methylene Chloride	5.098	84	18040	4.753	ug/l	93
43) Isopropyl Acetate	8.568	43	62748	6.867	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
Data File : VN069067.D
Acq On : 20 Oct 2021 17:27
Operator : JC/MD
Sample : M4252-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TEST-BA-(1)

Quant Time: Oct 20 23:15:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

