

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
 Data File : VN069068.D
 Acq On : 20 Oct 2021 17:52
 Operator : JC/MD
 Sample : M4252-04
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TEST-BB-(1)

Quant Time: Oct 20 23:15:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	310617	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	544002	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	529957	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	194857	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	204338	49.895	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.780%
35) Dibromofluoromethane	8.024	113	160094	50.582	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	101.160%
50) Toluene-d8	10.443	98	686525	54.211	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.420%
62) 4-Bromofluorobenzene	12.733	95	244377	51.147	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.300%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	11302	7.395	ug/l	98
18) Methyl Acetate	4.873	43	10126	1.747	ug/l	94
20) Methylene Chloride	5.106	84	17332	4.927	ug/l	90
43) Isopropyl Acetate	8.568	43	48021	5.671	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
Data File : VN069068.D
Acq On : 20 Oct 2021 17:52
Operator : JC/MD
Sample : M4252-04
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TEST-BB-(1)

Quant Time: Oct 20 23:15:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

