

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
 Data File : VN069069.D
 Acq On : 20 Oct 2021 18:17
 Operator : JC/MD
 Sample : M4252-07
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-CA-(1)

Quant Time: Oct 20 23:15:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	353675	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	614857	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	585210	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	216051	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.442	65	229675	49.254	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.500%
35) Dibromofluoromethane	8.024	113	180066	50.336	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.680%
50) Toluene-d8	10.446	98	766179	53.529	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	107.060%
62) 4-Bromofluorobenzene	12.733	95	270016	50.001	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.000%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	20178	11.595	ug/l	97
18) Methyl Acetate	4.864	43	10009	1.516	ug/l	98
20) Methylene Chloride	5.101	84	16940	4.229	ug/l	96
43) Isopropyl Acetate	8.571	43	49011	5.121	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
Data File : VN069069.D
Acq On : 20 Oct 2021 18:17
Operator : JC/MD
Sample : M4252-07
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MH-CA-(1)

Quant Time: Oct 20 23:15:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

