

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
 Data File : VN069071.D
 Acq On : 20 Oct 2021 19:07
 Operator : JC/MD
 Sample : M4260-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TEST-DA-(1)

Quant Time: Oct 20 23:15:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.091	168	340826	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	601460	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.746	117	580011	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.677	152	216366	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	221655	49.326	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.660%
35) Dibromofluoromethane	8.026	113	175680	50.204	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.400%
50) Toluene-d8	10.446	98	757045	54.069	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.140%
62) 4-Bromofluorobenzene	12.733	95	268926	50.908	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.820%
Target Compounds						
					Qvalue	
16) Acetone	4.296	43	11509	6.863	ug/l	93
18) Methyl Acetate	4.872	43	11430	1.797	ug/l	96
20) Methylene Chloride	5.106	84	18947	4.909	ug/l	94
43) Isopropyl Acetate	8.571	43	54811	5.855	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
Data File : VN069071.D
Acq On : 20 Oct 2021 19:07
Operator : JC/MD
Sample : M4260-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TEST-DA-(1)

Quant Time: Oct 20 23:15:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

