

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
 Data File : VN069052.D
 Acq On : 20 Oct 2021 10:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 20 16:37:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	0.546	0.562	-2.9	70	0.00
3 P	Chloromethane	0.611	0.577	5.6	67	0.00
4 C	Vinyl Chloride	0.621	0.693	-11.6#	77	0.00
5 T	Bromomethane	0.341	0.446	-30.8#	88	0.00
6 T	Chloroethane	0.366	0.448	-22.4	85	0.00
7 T	Trichlorofluoromethane	0.869	0.904	-4.0	74	0.00
8 T	Diethyl Ether	0.324	0.331	-2.2	73	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.467	0.507	-8.6	77	0.00
10 T	Methyl Iodide	0.678	0.637	6.0	67	0.00
11 T	Tert butyl alcohol	0.120	0.105	12.5	65	0.00
12 CM	1,1-Dichloroethene	0.483	0.469	2.9#	72	0.00
13 T	Acrolein	0.053	0.059	-11.3	92	0.00
14 T	Allyl chloride	0.866	0.837	3.3	67	0.00
15 T	Acrylonitrile	0.270	0.268	0.7	70	0.00
16 T	Acetone	0.246	0.299	-21.5	86	0.00
17 T	Carbon Disulfide	1.311	1.215	7.3	65	0.00
18 T	Methyl Acetate	0.933	0.954	-2.3	73	0.00
19 T	Methyl tert-butyl Ether	1.799	1.897	-5.4	74	0.00
20 T	Methylene Chloride	0.566	0.569	-0.5	74	0.00
21 T	trans-1,2-Dichloroethene	0.522	0.516	1.1	71	0.00
22 T	Diisopropyl ether	1.801	1.902	-5.6	72	0.00
23 T	Vinyl Acetate	1.362	1.409	-3.5	71	0.00
24 P	1,1-Dichloroethane	0.979	1.021	-4.3	74	0.00
25 T	2-Butanone	0.372	0.401	-7.8	74	0.00
26 T	2,2-Dichloropropane	0.933	0.964	-3.3	74	0.00
27 T	cis-1,2-Dichloroethene	0.631	0.637	-1.0	73	0.00
28 T	Bromochloromethane	0.443	0.439	0.9	74	0.00
29 T	Tetrahydrofuran	0.241	0.241	0.0	68	0.00
30 C	Chloroform	1.031	1.109	-7.6#	75	0.00
31 T	Cyclohexane	1.025	0.937	8.6	69	0.00
32 T	1,1,1-Trichloroethane	0.961	0.992	-3.2	73	0.00
33 S	1,2-Dichloroethane-d4	0.659	0.679	-3.0	84	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.291	0.292	-0.3	82	0.00
36 T	1,1-Dichloropropene	0.427	0.456	-6.8	74	0.00
37 T	Ethyl Acetate	0.441	0.444	-0.7	70	0.00
38 T	Carbon Tetrachloride	0.469	0.484	-3.2	72	0.00
39 T	Methylcyclohexane	0.550	0.583	-6.0	73	0.00
40 TM	Benzene	1.300	1.365	-5.0	73	0.00
41 T	Methacrylonitrile	0.222	0.233	-5.0	71	0.00
42 TM	1,2-Dichloroethane	0.461	0.511	-10.8	76	0.00
43 T	Isopropyl Acetate	0.778	0.801	-3.0	70	0.00
44 TM	Trichloroethene	0.325	0.338	-4.0	72	0.00
45 C	1,2-Dichloropropane	0.337	0.360	-6.8#	74	0.00
46 T	Dibromomethane	0.230	0.250	-8.7	76	0.00
47 T	Bromodichloromethane	0.477	0.512	-7.3	74	0.00
48 T	Methyl methacrylate	0.366	0.372	-1.6	70	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
 Data File : VN069052.D
 Acq On : 20 Oct 2021 10:40
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 20 16:37:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 13 05:44:40 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	69	0.00
50 S	Toluene-d8	1.164	1.191	-2.3	82	0.00
51 T	4-Methyl-2-Pentanone	0.451	0.467	-3.5	70	0.00
52 CM	Toluene	0.837	0.904	-8.0#	73	0.00
53 T	t-1,3-Dichloropropene	0.541	0.588	-8.7	75	0.00
54 T	cis-1,3-Dichloropropene	0.563	0.608	-8.0	74	0.00
55 T	1,1,2-Trichloroethane	0.330	0.360	-9.1	76	0.00
56 T	Ethyl methacrylate	0.554	0.595	-7.4	74	0.00
57 T	1,3-Dichloropropane	0.544	0.609	-11.9	76	0.00
58 T	2-Chloroethyl Vinyl ether	0.134	0.142	-6.0	73	0.00
59 T	2-Hexanone	0.324	0.354	-9.3	74	0.00
60 T	Dibromochloromethane	0.372	0.396	-6.5	74	0.00
61 T	1,2-Dibromoethane	0.338	0.373	-10.4	75	0.00
62 S	4-Bromofluorobenzene	0.439	0.469	-6.8	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	78	0.00
64 T	Tetrachloroethene	0.322	0.305	5.3	67	0.00
65 PM	Chlorobenzene	0.946	1.031	-9.0	78	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.386	-4.3	74	0.00
67 C	Ethyl Benzene	1.758	1.904	-8.3#	76	0.00
68 T	m/p-Xylenes	0.666	0.724	-8.7	76	0.00
69 T	o-Xylene	0.667	0.717	-7.5	76	0.00
70 T	Styrene	1.086	1.208	-11.2	78	0.00
71 P	Bromoform	0.297	0.291	2.0	69	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
73 T	Isopropylbenzene	4.162	4.264	-2.5	77	0.00
74 T	N-amyl acetate	1.577	1.611	-2.2	75	0.00
75 P	1,1,2,2-Tetrachloroethane	1.203	1.265	-5.2	79	0.00
76 T	1,2,3-Trichloropropane	1.064	0.972	8.6	64	0.00
77 T	Bromobenzene	0.920	0.944	-2.6	77	0.00
78 T	n-propylbenzene	4.677	4.983	-6.5	79	0.00
79 T	2-Chlorotoluene	2.883	3.020	-4.8	79	0.00
80 T	1,3,5-Trimethylbenzene	3.466	3.565	-2.9	77	0.00
81 T	trans-1,4-Dichloro-2-butene	0.445	0.421	5.4	73	0.00
82 T	4-Chlorotoluene	2.786	2.998	-7.6	80	0.00
83 T	tert-Butylbenzene	3.018	3.094	-2.5	77	0.00
84 T	1,2,4-Trimethylbenzene	3.340	3.506	-5.0	78	0.00
85 T	sec-Butylbenzene	4.179	4.470	-7.0	79	0.00
86 T	p-Isopropyltoluene	3.352	3.624	-8.1	80	0.00
87 T	1,3-Dichlorobenzene	1.592	1.702	-6.9	79	0.00
88 T	1,4-Dichlorobenzene	1.580	1.683	-6.5	79	0.00
89 T	n-Butylbenzene	2.791	3.099	-11.0	81	0.00
90 T	Hexachloroethane	0.704	0.695	1.3	74	0.00
91 T	1,2-Dichlorobenzene	1.546	1.629	-5.4	77	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.232	0.228	1.7	71	0.00
93 T	1,2,4-Trichlorobenzene	0.686	0.779	-13.6	78	0.00
94 T	Hexachlorobutadiene	0.440	0.423	3.9	71	0.00
95 T	Naphthalene	1.793	2.100	-17.1	78	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102021\
Data File : VN069052.D
Acq On : 20 Oct 2021 10:40
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Oct 20 16:37:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N101221W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 13 05:44:40 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.650	0.742	-14.2	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6