

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102023\
 Data File : VN079622.D
 Acq On : 20 Oct 2023 12:29
 Operator : JC\MD
 Sample : VN1020WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1020WBL01

Quant Time: Oct 20 23:48:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101723W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Oct 18 02:19:15 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	18967	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	127463	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	148558	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	82745	30.684	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	102.267%
60) 4-Bromofluorobenzene	12.853	95	56038	20.627	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	68.767%
63) Toluene-d8	10.571	98	183317	28.797	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102023\
Data File : VN079622.D
Acq On : 20 Oct 2023 12:29
Operator : JC\MD
Sample : VN1020WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1020WBL01

Quant Time: Oct 20 23:48:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N101723W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Wed Oct 18 02:19:15 2023
Response via : Initial Calibration

