

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102025\
 Data File : VN088073.D
 Acq On : 20 Oct 2025 09:26
 Operator : JC\MD
 Sample : VN1020WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Oct 21 05:35:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100825W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Oct 09 02:36:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	62285	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.082	114	343376	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.847	117	304684	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	156635	32.296	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	= 107.667%	
60) 4-Bromofluorobenzene	12.829	95	152913	30.443	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	= 101.467%	
63) Toluene-d8	10.547	98	420133	30.949	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	= 103.167%	
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.142	85	77653	21.767	ug/l	92
3) Chloromethane	2.383	50	76792	18.986	ug/l	95
4) Vinyl Chloride	2.536	62	85262	20.217	ug/l	97
5) Bromomethane	2.983	94	50875	22.434	ug/l	98
6) Chloroethane	3.136	64	60988	21.923	ug/l	95
7) Trichlorofluoromethane	3.512	101	130211	22.705	ug/l	99
8) Diethyl Ether	3.959	74	49392	19.061	ug/l	97
9) 1,1,2-Trichlorotrifluo...	4.365	101	67557	22.006	ug/l	96
10) 1,1-Dichloroethene	4.342	96	68131	20.804	ug/l	95
11) Methyl Iodide	4.589	142	52620	17.434	ug/l	93
12) Methyl Acetate	5.018	43	116619	21.905	ug/l	93
13) Acrolein	4.183	56	59308	71.986	ug/l #	4
14) Acrylonitrile	5.706	53	254622	99.104	ug/l	100
15) Acetone	4.424	58	67752	90.045	ug/l	87
16) Carbon Disulfide	4.706	76	202681	19.743	ug/l	96
17) Allyl chloride	5.018	41	118111	19.113	ug/l	95
18) Methylene Chloride	5.265	84	79795	19.171	ug/l	96
19) trans-1,2-Dichloroethene	5.777	96	74543	19.687	ug/l	89
20) Diisopropyl ether	6.659	45	279278	19.851	ug/l	99
21) 1,1-Dichloroethane	6.553	63	147591	20.152	ug/l	97
22) cis-1,2-Dichloroethene	7.471	96	91891	19.387	ug/l	96
23) tert-Butyl Alcohol	5.518	59	95731	92.904	ug/l #	100
24) Methyl tert-Butyl Ether	5.783	73	275566	19.494	ug/l #	78
25) Chloroform	7.947	83	150557	19.921	ug/l	97
26) Cyclohexane	8.241	56	125784	20.729	ug/l #	95
29) 1,1-Dichloropropene	8.353	75	104611	21.254	ug/l	97
30) 2-Butanone	7.471	43	362293	99.803	ug/l	98
31) 2,2-Dichloropropane	7.471	77	134132	21.150	ug/l	98
32) 1,1,1-Trichloroethane	8.147	97	130492	21.218	ug/l	98
33) Carbon Tetrachloride	8.341	117	110825	21.513	ug/l	98
34) Benzene	8.588	78	323530	19.817	ug/l	98
35) Methacrylonitrile	7.765	41	79995	20.956	ug/l	98
36) 1,2-Dichloroethane	8.653	62	120255	21.218	ug/l	97
37) Trichloroethene	9.329	130	75973	19.642	ug/l	97
38) Methylcyclohexane	9.577	83	126136	20.752	ug/l	96
39) 1,2-Dichloropropane	9.600	63	82623	20.184	ug/l	100
40) Dibromomethane	9.688	93	61481	20.633	ug/l	94
41) Bromodichloromethane	9.865	83	124256	20.418	ug/l	98
42) Vinyl Acetate	6.588	43	1170698	94.666	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102025\
 Data File : VN088073.D
 Acq On : 20 Oct 2025 09:26
 Operator : JC\MD
 Sample : VN1020WBS01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1


Instrument :
 MSVOA_N
 ClientSampleId :


Quant Time: Oct 21 05:35:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100825W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Oct 09 02:36:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
43) Ethyl Acetate	7.547	43	139785	20.032	ug/l	98
44) Isopropyl Acetate	8.671	43	224369	19.686	ug/l	98
45) 1,4-Dioxane	9.677	88	29977	395.465	ug/l #	94
46) Methyl methacrylate	9.665	41	104697	20.859	ug/l	94
48) t-1,3-Dichloropropene	10.812	75	133458	19.721	ug/l	98
49) cis-1,3-Dichloropropene	10.288	75	135218	19.155	ug/l	92
50) 1,1,2-Trichloroethane	10.994	97	78759	19.959	ug/l	95
51) Ethyl methacrylate	10.859	69	129959	19.460	ug/l	98
52) 1,3-Dichloropropane	11.141	76	137735	20.006	ug/l	100
53) Dibromochloromethane	11.335	129	91345	19.363	ug/l	100
54) 1,2-Dibromoethane	11.447	107	80700	19.129	ug/l	100
55) 2-Chloroethyl vinyl ether	10.141	63	358393	99.423	ug/l	98
56) Bromoform	12.559	173	57617	18.065	ug/l	98
58) 4-Methyl-2-Pentanone	10.424	43	689865	101.227	ug/l	98
59) 2-Hexanone	11.176	43	439800	97.743	ug/l	98
61) Tetrachloroethene	11.082	164	59740	19.362	ug/l	95
62) Toluene	10.606	91	346319	20.038	ug/l	100
64) Chlorobenzene	11.871	112	216235	19.588	ug/l	100
65) 1,1,1,2-Tetrachloroethane	11.941	131	76374	20.076	ug/l	99
66) Ethyl Benzene	11.941	91	382033	20.329	ug/l	98
67) m/p-Xylenes	12.047	106	288899	39.877	ug/l	97
68) o-Xylene	12.376	106	139692	19.369	ug/l	96
69) Styrene	12.394	104	235197	19.714	ug/l	99
70) Isopropylbenzene	12.676	105	361962	20.478	ug/l	98
71) 1,1,2,2-Tetrachloroethane	12.918	83	120825	20.162	ug/l	99
72) 1,2,3-Trichloropropane	12.970	75	158897	28.550	ug/l	58
73) Bromobenzene	12.959	156	79169	18.325	ug/l	88
74) n-propylbenzene	13.012	91	445643	20.885	ug/l	97
75) 2-Chlorotoluene	13.106	91	263437	20.191	ug/l	97
76) 1,3,5-Trimethylbenzene	13.153	105	306719	20.428	ug/l	96
77) t-1,4-Dichloro-2-butene	12.718	75	40290	18.165	ug/l	95
78) 4-Chlorotoluene	13.200	91	266609	20.113	ug/l	98
79) tert-butylbenzene	13.417	119	268223	20.607	ug/l	96
80) 1,2,4-Trimethylbenzene	13.459	105	302874	20.150	ug/l	97
81) sec-Butylbenzene	13.594	105	391552	21.395	ug/l	97
82) p-Isopropyltoluene	13.706	119	329248	21.072	ug/l	99
83) 1,3-Dichlorobenzene	13.712	146	157162	19.334	ug/l	98
84) 1,4-Dichlorobenzene	13.794	146	163893	19.550	ug/l	99
85) n-Butylbenzene	14.035	91	316165	21.750	ug/l	97
86) Hexachloroethane	14.312	117	66588	21.274	ug/l	85
87) 1,2-Dichlorobenzene	14.088	146	152656	19.093	ug/l	99
88) 1,2-Dibromo-3-Chloropr...	14.700	75	29509	21.440	ug/l	87
89) 1,2,4-Trichlorobenzene	15.811	180	87549	18.383	ug/l	98
90) Hexachlorobutadiene	15.470	225	33672	20.315	ug/l	95
91) Naphthalene	15.617	128	326635	18.734	ug/l	99
92) 1,2,3-Trichlorobenzene	15.811	180	87549	18.383	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Oct 21 05:35:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100825W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Oct 09 02:36:32 2025
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 7.800 min Scan# 993
Delta R.T. 0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:128 Resp: 62285
Ion Ratio Lower Upper
128 100
49 184.7 0.0 439.8
130 130.1 0.0 325.8



#2
Dichlorodifluoromethane
Concen: 21.767 ug/l
RT: 2.142 min Scan# 32
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 85 Resp: 77653
Ion Ratio Lower Upper
85 100
87 31.0 17.9 53.8









#3
Chloromethane
Concen: 18.986 ug/l
RT: 2.383 min Scan# 73
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 50 Resp: 76792
Ion Ratio Lower Upper
50 100
52 35.7 26.3 39.5



#4
Vinyl Chloride
Concen: 20.217 ug/l
RT: 2.536 min Scan# 99
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26


Tgt Ion: 62 Resp: 85262
Ion Ratio Lower Upper
62 100
64 30.3 25.7 38.5








#5
Bromomethane
Concen: 22.434 ug/l
RT: 2.983 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 94 Resp: 50875
Ion Ratio Lower Upper
94 100
96 92.9 73.0 109.4



#6
Chloroethane
Concen: 21.923 ug/l
RT: 3.136 min Scan# 201
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 64 Resp: 60988
Ion Ratio Lower Upper
64 100
66 30.5 26.8 40.2









#7
 Trichlorofluoromethane
 Concen: 22.705 ug/l
 RT: 3.512 min Scan# 20
 Delta R.T. 0.006 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:101 Resp: 130211
 Ion Ratio Lower Upper
 101 100
 103 64.6 51.0 76.4



#8
 Diethyl Ether
 Concen: 19.061 ug/l
 RT: 3.959 min Scan# 341
 Delta R.T. 0.000 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26

Tgt Ion: 74 Resp: 49392
 Ion Ratio Lower Upper
 74 100
 45 105.2 20.4 183.8









#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 22.006 ug/l
 RT: 4.365 min Scan# 410
 Delta R.T. 0.000 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:101 Resp: 67557
 Ion Ratio Lower Upper
 101 100
 85 47.6 0.0 103.2
 151 68.3 0.0 141.4



#10
 1,1-Dichloroethene
 Concen: 20.804 ug/l
 RT: 4.342 min Scan# 406
 Delta R.T. 0.012 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26


Tgt Ion: 96 Resp: 68131
 Ion Ratio Lower Upper
 96 100
 61 170.8 144.1 216.1
 98 61.2 49.0 73.6








#11
Methyl Iodide
Concen: 17.434 ug/l
RT: 4.589 min Scan# 446
Delta R.T. 0.012 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:142 Resp: 52620
Ion Ratio Lower Upper
142 100
127 43.4 23.4 70.3
141 11.6 8.8 26.2



#12
Methyl Acetate
Concen: 21.905 ug/l
RT: 5.018 min Scan# 521
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 43 Resp: 116619
Ion Ratio Lower Upper
43 100
74 25.4 23.3 34.9









#13
Acrolein
Concen: 71.986 ug/l
RT: 4.183 min Scan# 31
Delta R.T. 0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 56 Resp: 59308
Ion Ratio Lower Upper
56 100
55 73.0 0.3 0.5#



#14
Acrylonitrile
Concen: 99.104 ug/l
RT: 5.706 min Scan# 638
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 53 Resp: 254622
Ion Ratio Lower Upper
53 100
52 82.4 41.3 124.1
51 36.1 18.0 54.0









#15
Acetone
Concen: 90.045 ug/l
RT: 4.424 min Scan# 41
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 58 Resp: 67752
Ion Ratio Lower Upper
58 100
43 325.0 282.6 424.0



#16
Carbon Disulfide
Concen: 19.743 ug/l
RT: 4.706 min Scan# 468
Delta R.T. 0.012 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26


Tgt Ion: 76 Resp: 202681
Ion Ratio Lower Upper
76 100
78 8.3 7.8 11.8








#17
 Allyl chloride
 Concen: 19.113 ug/l
 RT: 5.018 min Scan# 51
 Delta R.T. 0.006 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion: 41 Resp: 118111
 Ion Ratio Lower Upper
 41 100
 39 76.0 36.4 109.2
 76 41.2 18.1 54.1



#18
 Methylene Chloride
 Concen: 19.171 ug/l
 RT: 5.265 min Scan# 563
 Delta R.T. 0.006 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26

Tgt Ion: 84 Resp: 79795
 Ion Ratio Lower Upper
 84 100
 49 125.8 102.6 153.8
 51 39.8 28.6 42.8
 86 55.7 48.2 72.2









#19
trans-1,2-Dichloroethene
Concen: 19.687 ug/l
RT: 5.777 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 96 Resp: 74543
Ion Ratio Lower Upper
96 100
61 141.3 100.6 151.0
98 62.0 54.0 81.0



#20
Diisopropyl ether
Concen: 19.851 ug/l
RT: 6.659 min Scan# 800
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26


Tgt Ion: 45 Resp: 279278
Ion Ratio Lower Upper
45 100
43 57.5 46.1 69.1
87 26.3 22.3 33.5
59 11.3 8.6 13.0








#21
 1,1-Dichloroethane
 Concen: 20.152 ug/l
 RT: 6.553 min Scan# 71
 Delta R.T. 0.000 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion: 63 Resp: 147591
 Ion Ratio Lower Upper
 63 100
 98 7.4 4.3 12.8
 100 4.2 2.6 8.0



#22
 cis-1,2-Dichloroethene
 Concen: 19.387 ug/l
 RT: 7.471 min Scan# 938
 Delta R.T. 0.000 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26


Tgt Ion: 96 Resp: 91891
 Ion Ratio Lower Upper
 96 100
 61 152.5 117.0 175.6
 98 65.1 52.6 79.0








#23
tert-Butyl Alcohol
Concen: 92.904 ug/l
RT: 5.518 min Scan# 608
Delta R.T. -0.012 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion: 59 Resp: 95731
Ion Ratio Lower Upper
59 100
52 0.0 0.0 0.0



#24
Methyl tert-Butyl Ether
Concen: 19.494 ug/l
RT: 5.783 min Scan# 651
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 73 Resp: 275566
Ion Ratio Lower Upper
73 100
43 10.7 17.0 25.4#





#25

Chloroform

Concen: 19.921 ug/l

RT: 7.947 min Scan# 1019

Delta R.T. -0.006 min

Lab File: VN088073.D

Acq: 20 Oct 2025 09:26

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 83 Resp: 150557

Ion Ratio Lower Upper

83 100

85 61.3 50.9 76.3



#26

Cyclohexane

Concen: 20.729 ug/l

RT: 8.241 min Scan# 1069

Delta R.T. 0.000 min

Lab File: VN088073.D

Acq: 20 Oct 2025 09:26




Tgt Ion: 56 Resp: 125784


Ion Ratio Lower Upper

56 100

89 27.6 8.6 12.8#

84 87.0 69.6 104.4





#27

1,2-Dichloroethane-d4

Concen: 32.296 ug/l

RT: 8.559 min Scan# 1123

Delta R.T. -0.006 min

Lab File: VN088073.D

Acq: 20 Oct 2025 09:26

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 65 Resp: 156635

Ion Ratio Lower Upper

65 100

67 52.0 41.9 62.9



#28

1,4-Difluorobenzene

Concen: 30.000 ug/l

RT: 9.082 min Scan# 1212

Delta R.T. 0.000 min

Lab File: VN088073.D

Acq: 20 Oct 2025 09:26




Tgt Ion: 114 Resp: 343376


Ion Ratio Lower Upper

114 100

63 22.0 16.6 25.0

88 16.2 12.6 18.8









#29
1,1-Dichloropropene
Concen: 21.254 ug/l
RT: 8.353 min Scan# 1088
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 104611
Ion Ratio Lower Upper
75 100
110 33.6 0.0 72.2
77 31.4 0.0 60.6



#30
2-Butanone
Concen: 99.803 ug/l
RT: 7.471 min Scan# 938
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 43 Resp: 362293
Ion Ratio Lower Upper
43 100
57 8.5 6.5 9.7
72 25.8 21.4 32.2









#31
2,2-Dichloropropane
Concen: 21.150 ug/l
RT: 7.471 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 77 Resp: 134132
Ion Ratio Lower Upper
77 100
97 21.5 17.8 26.8



#32
1,1,1-Trichloroethane
Concen: 21.218 ug/l
RT: 8.147 min Scan# 1053
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26


Tgt Ion: 97 Resp: 130492
Ion Ratio Lower Upper
97 100
99 63.4 51.4 77.2
61 45.3 34.7 52.1








#33
Carbon Tetrachloride
Concen: 21.513 ug/l
RT: 8.341 min Scan# 1108
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Instrument :
MSVOA_N
ClientSampleId :







Tgt Ion:117 Resp: 110825
Ion Ratio Lower Upper
117 100
119 99.0 77.9 116.9
121 33.8 26.4 39.6



#34
Benzene
Concen: 19.817 ug/l
RT: 8.588 min Scan# 1128
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 78 Resp: 323530
Ion Ratio Lower Upper
78 100
77 24.7 18.9 28.3
51 20.0 15.6 23.4









#35
Methacrylonitrile
Concen: 20.956 ug/l
RT: 7.765 min Scan# 91
Delta R.T. 0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:	41	Resp:	79995
Ion	Ratio	Lower	Upper
41	100		
39	53.4	26.1	78.3
67	66.3	33.1	99.2
52	27.7	15.4	46.1



#36
1,2-Dichloroethane
Concen: 21.218 ug/l
RT: 8.653 min Scan# 1139
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion:	62	Resp:	120255
Ion	Ratio	Lower	Upper
62	100		
98	8.6	7.8	11.6









#37
Trichloroethene
Concen: 19.642 ug/l
RT: 9.329 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:130 Resp: 75973
Ion Ratio Lower Upper
130 100
95 101.0 78.7 118.1



#38
Methylcyclohexane
Concen: 20.752 ug/l
RT: 9.577 min Scan# 1296
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 83 Resp: 126136
Ion Ratio Lower Upper
83 100
55 79.1 37.3 111.9
98 49.2 25.4 76.0





#39

1,2-Dichloropropane

Concen: 20.184 ug/l

RT: 9.600 min Scan# 1300

Delta R.T. 0.000 min

Lab File: VN088073.D

Acq: 20 Oct 2025 09:26

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion: 63 Resp: 82623

Ion Ratio Lower Upper

63 100

65 31.3 25.0 37.6



#40

Dibromomethane

Concen: 20.633 ug/l

RT: 9.688 min Scan# 1315

Delta R.T. 0.000 min

Lab File: VN088073.D

Acq: 20 Oct 2025 09:26




Tgt Ion: 93 Resp: 61481


Ion Ratio Lower Upper

93 100

95 82.0 0.0 158.6

174 82.6 0.0 180.8









#41
Bromodichloromethane
Concen: 20.418 ug/l
RT: 9.865 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 83 Resp: 124256
Ion Ratio Lower Upper
83 100
85 65.1 51.0 76.4
127 7.9 6.1 9.1



#42
Vinyl Acetate
Concen: 94.666 ug/l
RT: 6.588 min Scan# 788
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 43 Resp: 1170698
Ion Ratio Lower Upper
43 100
86 9.7 8.4 12.6









#43
Ethyl Acetate
Concen: 20.032 ug/l
RT: 7.547 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 139785
Ion Ratio Lower Upper
43 100
45 13.7 10.2 15.4



#44
Isopropyl Acetate
Concen: 19.686 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 43 Resp: 224369
Ion Ratio Lower Upper
43 100
61 24.6 18.7 28.1





#45

1,4-Dioxane

Concen: 395.465 ug/l

RT: 9.677 min Scan# 11

Delta R.T. 0.000 min

Lab File: VN088073.D

Acq: 20 Oct 2025 09:26

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 88 Resp: 29977

Ion Ratio Lower Upper

88 100

43 28.7 18.6 27.8#

58 75.4 57.4 86.2



#46

Methyl methacrylate

Concen: 20.859 ug/l

RT: 9.665 min Scan# 1311

Delta R.T. 0.000 min

Lab File: VN088073.D

Acq: 20 Oct 2025 09:26




Tgt Ion: 41 Resp: 104697


Ion Ratio Lower Upper

41 100

69 81.0 44.9 134.7

39 61.5 30.3 90.8









#48
 t-1,3-Dichloropropene
 Concen: 19.721 ug/l
 RT: 10.812 min Scan# 11
 Delta R.T. -0.006 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion: 75 Resp: 133458
 Ion Ratio Lower Upper
 75 100
 77 29.7 24.5 36.7



#49
 cis-1,3-Dichloropropene
 Concen: 19.155 ug/l
 RT: 10.288 min Scan# 1417
 Delta R.T. -0.006 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26


Tgt Ion: 75 Resp: 135218
 Ion Ratio Lower Upper
 75 100
 77 32.1 25.0 37.4
 39 54.4 37.4 56.0








#50
 1,1,2-Trichloroethane
 Concen: 19.959 ug/l
 RT: 10.994 min Scan# 1537
 Delta R.T. 0.000 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26

Instrument :
 MSVOA_N
 ClientSampleId :







Tgt Ion:	97	Resp:	78759
Ion	Ratio	Lower	Upper
97	100		
83	83.0	72.5	108.7
85	55.7	45.5	68.3
99	66.6	50.0	75.0



#51
 Ethyl methacrylate
 Concen: 19.460 ug/l
 RT: 10.859 min Scan# 1514
 Delta R.T. 0.000 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26

Tgt Ion:	69	Resp:	129959
Ion	Ratio	Lower	Upper
69	100		
41	64.6	33.3	99.8
39	39.8	20.3	61.0









#52
1,3-Dichloropropane
Concen: 20.006 ug/l
RT: 11.141 min Scan# 1562
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 76 Resp: 137735
Ion Ratio Lower Upper
76 100
78 32.4 0.0 65.0



#53
Dibromochloromethane
Concen: 19.363 ug/l
RT: 11.335 min Scan# 1595
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 129 Resp: 91345
Ion Ratio Lower Upper
129 100
127 78.0 39.1 117.3









#54
1,2-Dibromoethane
Concen: 19.129 ug/l
RT: 11.447 min Scan# 1614
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:107 Resp: 80700
Ion Ratio Lower Upper
107 100
109 94.9 75.5 113.3



#55
2-Chloroethyl vinyl ether
Concen: 99.423 ug/l
RT: 10.141 min Scan# 1392
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 63 Resp: 358393
Ion Ratio Lower Upper
63 100
65 32.4 25.7 38.5
106 26.0 22.6 34.0









#56
Bromoform
Concen: 18.065 ug/l
RT: 12.559 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:173 Resp: 57617
Ion Ratio Lower Upper
173 100
175 48.5 39.6 59.4
254 7.1 6.6 10.0



#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 11.847 min Scan# 1682
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion:117 Resp: 304684
Ion Ratio Lower Upper
117 100
82 58.5 45.2 67.8
119 31.9 25.9 38.9









#58
4-Methyl-2-Pentanone
Concen: 101.227 ug/l
RT: 10.424 min Scan# 1440
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 43 Resp: 689865
Ion Ratio Lower Upper
43 100
58 37.6 31.2 46.8
85 16.2 13.7 20.5



#59
2-Hexanone
Concen: 97.743 ug/l
RT: 11.176 min Scan# 1568
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 43 Resp: 439800
Ion Ratio Lower Upper
43 100
58 55.2 43.1 64.7
57 18.6 14.6 21.8









#60
4-Bromofluorobenzene
Concen: 30.443 ug/l
RT: 12.829 min Scan# 1849
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 95 Resp: 152913
Ion Ratio Lower Upper
95 100
174 65.5 57.4 86.0
176 64.9 57.0 85.4



#61
Tetrachloroethene
Concen: 19.362 ug/l
RT: 11.082 min Scan# 1552
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion:164 Resp: 59740
Ion Ratio Lower Upper
164 100
166 126.5 97.2 145.8
129 94.9 73.7 110.5
131 97.2 71.5 107.3









#62
Toluene
Concen: 20.038 ug/l
RT: 10.606 min Scan# 1472
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 346319
Ion Ratio Lower Upper
91 100
92 57.0 45.8 68.8



#63
Toluene-d8
Concen: 30.949 ug/l
RT: 10.547 min Scan# 1461
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 98 Resp: 420133
Ion Ratio Lower Upper
98 100
100 65.3 53.1 79.7





#64

Chlorobenzene

Concen: 19.588 ug/l

RT: 11.871 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN088073.D

Acq: 20 Oct 2025 09:26

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:112 Resp: 216235

Ion Ratio Lower Upper

112 100

114 31.9 25.7 38.5



#65

1,1,1,2-Tetrachloroethane

Concen: 20.076 ug/l

RT: 11.941 min Scan# 1698

Delta R.T. 0.000 min

Lab File: VN088073.D

Acq: 20 Oct 2025 09:26




Tgt Ion:131 Resp: 76374


Ion Ratio Lower Upper

131 100

133 93.6 0.0 183.8

119 65.4 0.0 132.0









#66
Ethyl Benzene
Concen: 20.329 ug/l
RT: 11.941 min Scan# 1698
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 382033
Ion Ratio Lower Upper
91 100
106 31.3 24.0 36.0



#67
m/p-Xylenes
Concen: 39.877 ug/l
RT: 12.047 min Scan# 1716
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion:106 Resp: 288899
Ion Ratio Lower Upper
106 100
91 207.0 161.4 242.0









#68
o-Xylene
Concen: 19.369 ug/l
RT: 12.376 min Scan# 1772
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:106 Resp: 139692
Ion Ratio Lower Upper
106 100
91 216.3 105.2 315.6



#69
Styrene
Concen: 19.714 ug/l
RT: 12.394 min Scan# 1775
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion:104 Resp: 235197
Ion Ratio Lower Upper
104 100
78 52.0 41.8 62.6
103 54.3 44.9 67.3









#70
Isopropylbenzene
Concen: 20.478 ug/l
RT: 12.676 min Scan# 1823
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:105 Resp: 361962
Ion Ratio Lower Upper
105 100
120 26.1 19.1 35.5



#71
1,1,2,2-Tetrachloroethane
Concen: 20.162 ug/l
RT: 12.918 min Scan# 1864
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 83 Resp: 120825
Ion Ratio Lower Upper
83 100
131 8.7 4.8 14.3
85 64.3 32.4 97.2









#72
1,2,3-Trichloropropane
Concen: 28.550 ug/l
RT: 12.970 min Scan# 1874
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 75 Resp: 158897
Ion Ratio Lower Upper
75 100
77 121.5 0.0 364.0



#73
Bromobenzene
Concen: 18.325 ug/l
RT: 12.959 min Scan# 1871
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion:156 Resp: 79169
Ion Ratio Lower Upper
156 100
77 243.8 0.0 439.8
158 99.4 0.0 190.0









#74
n-propylbenzene
Concen: 20.885 ug/l
RT: 13.012 min Scan# 1880
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 445643
Ion Ratio Lower Upper
91 100
120 21.9 0.0 46.2



#75
2-Chlorotoluene
Concen: 20.191 ug/l
RT: 13.106 min Scan# 1896
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 91 Resp: 263437
Ion Ratio Lower Upper
91 100
126 31.5 0.0 66.6









#76
1,3,5-Trimethylbenzene
Concen: 20.428 ug/l
RT: 13.153 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:105 Resp: 306719
Ion Ratio Lower Upper
105 100
120 47.7 0.0 100.6



#77
t-1,4-Dichloro-2-butene
Concen: 18.165 ug/l
RT: 12.718 min Scan# 1830
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 75 Resp: 40290
Ion Ratio Lower Upper
75 100
53 90.3 41.8 125.3
89 39.8 19.9 59.6









#78
4-Chlorotoluene
Concen: 20.113 ug/l
RT: 13.200 min Scan# 1912
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion: 91 Resp: 266609
Ion Ratio Lower Upper
91 100
126 33.5 0.0 68.8



#79
tert-butylbenzene
Concen: 20.607 ug/l
RT: 13.417 min Scan# 1949
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion: 119 Resp: 268223
Ion Ratio Lower Upper
119 100
91 70.0 0.0 132.4
134 24.5 0.0 48.6









#80
1,2,4-Trimethylbenzene
Concen: 20.150 ug/l
RT: 13.459 min Scan# 1956
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:105 Resp: 302874
Ion Ratio Lower Upper
105 100
120 44.7 0.0 93.8



#81
sec-Butylbenzene
Concen: 21.395 ug/l
RT: 13.594 min Scan# 1979
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26

Tgt Ion:105 Resp: 391552
Ion Ratio Lower Upper
105 100
134 19.0 0.0 41.2









#82
 p-Isopropyltoluene
 Concen: 21.072 ug/l
 RT: 13.706 min Scan# 1998
 Delta R.T. 0.000 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26




Instrument :
 MSVOA_N
 ClientSampleId :


Tgt Ion:119 Resp: 329248
 Ion Ratio Lower Upper
 119 100
 134 25.5 0.0 52.6
 91 24.0 0.0 48.2



#83
 1,3-Dichlorobenzene
 Concen: 19.334 ug/l
 RT: 13.712 min Scan# 1999
 Delta R.T. 0.000 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26


Tgt Ion:146 Resp: 157162
 Ion Ratio Lower Upper
 146 100
 111 45.1 21.1 63.1
 148 63.6 31.5 94.5








#84
1,4-Dichlorobenzene
Concen: 19.550 ug/l
RT: 13.794 min Scan# 2012
Delta R.T. 0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26


Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:146 Resp: 163893
Ion Ratio Lower Upper
146 100
111 42.6 20.5 61.6
148 63.2 31.9 95.9




#85
n-Butylbenzene
Concen: 21.750 ug/l
RT: 14.035 min Scan# 2054
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26



Tgt Ion: 91 Resp: 316165
Ion Ratio Lower Upper
91 100
92 52.4 0.0 107.6
134 23.7 0.0 52.6





#86

Hexachloroethane

Concen: 21.274 ug/l

RT: 14.312 min Scan# 2101

Delta R.T. 0.000 min

Lab File: VN088073.D

Acq: 20 Oct 2025 09:26

Instrument :

MSVOA_N





ClientSampleId :

Tgt Ion:117 Resp: 66588

Ion Ratio Lower Upper

117 100

201 57.4 34.8 104.3



#87

1,2-Dichlorobenzene

Concen: 19.093 ug/l

RT: 14.088 min Scan# 2063

Delta R.T. 0.000 min

Lab File: VN088073.D

Acq: 20 Oct 2025 09:26




Tgt Ion:146 Resp: 152656


Ion Ratio Lower Upper

146 100

111 45.7 22.3 66.8

148 62.7 31.4 94.3





#88

1,2-Dibromo-3-Chloropropane

Concen: 21.440 ug/l

RT: 14.700 min Scan# 2167

Delta R.T. 0.000 min

Lab File: VN088073.D

Acq: 20 Oct 2025 09:26

Instrument :

MSVOA_N

ClientSampleId :





Tgt Ion: 75 Resp: 29509

Ion Ratio Lower Upper

75 100

155 67.4 63.5 95.3

157 90.0 81.7 122.5



#89

1,2,4-Trichlorobenzene

Concen: 18.383 ug/l

RT: 15.811 min Scan# 2356

Delta R.T. 0.000 min

Lab File: VN088073.D

Acq: 20 Oct 2025 09:26




Tgt Ion:180 Resp: 87549


Ion Ratio Lower Upper

180 100

182 94.0 74.6 112.0

145 36.4 27.2 40.8









#90
Hexachlorobutadiene
Concen: 20.315 ug/l
RT: 15.470 min Scan# 21
Delta R.T. -0.006 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26




Instrument :
MSVOA_N
ClientSampleId :


Tgt Ion:225 Resp: 33672
Ion Ratio Lower Upper
225 100
223 61.0 0.0 129.6
227 62.0 0.0 132.2



#91
Naphthalene
Concen: 18.734 ug/l
RT: 15.617 min Scan# 2323
Delta R.T. 0.000 min
Lab File: VN088073.D
Acq: 20 Oct 2025 09:26


Tgt Ion:128 Resp: 326635
Ion Ratio Lower Upper
128 100
127 13.2 10.3 15.5
129 11.0 8.6 13.0





#92
 1,2,3-Trichlorobenzene
 Concen: 18.383 ug/l
 RT: 15.811 min Scan# 21
 Delta R.T. 0.000 min
 Lab File: VN088073.D
 Acq: 20 Oct 2025 09:26

Instrument :
 MSVOA_N
 ClientSampleId :



Tgt Ion:	180	Resp:	87549
Ion Ratio	Lower	Upper	
180	100		
182	94.0	0.0	186.6
145	36.4	0.0	68.0

