

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102025\
 Data File : VN088076.D
 Acq On : 20 Oct 2025 11:31
 Operator : JC\MD
 Sample : Q3385-01 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EFF-WW

Quant Time: Oct 21 05:37:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100825W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Oct 09 02:36:32 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.800	128	63544	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.083	114	356820	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.841	117	313448	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.559	65	159806	32.297	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	107.667%
60) 4-Bromofluorobenzene	12.829	95	148887	28.813	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	96.033%
63) Toluene-d8	10.547	98	433757	31.059	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.533%
Target Compounds						
					Qvalue	
15) Acetone	4.418	58	587860	765.813	ug/l	95
25) Chloroform	7.947	83	123897	16.068	ug/l	95
41) Bromodichloromethane	9.865	83	8427	1.333	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN102025\
Data File : VN088076.D
Acq On : 20 Oct 2025 11:31
Operator : JC\MD
Sample : Q3385-01 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EFF-WW

Quant Time: Oct 21 05:37:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N100825W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Oct 09 02:36:32 2025
Response via : Initial Calibration

