

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102119\
 Data File : VN058886.D
 Acq On : 21 Oct 2019 14:23
 Operator : JC/SP
 Sample : K5453-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RT1496

Quant Time: Oct 22 04:27:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	422612	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	682659	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	625898	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	243578	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	306977	49.53	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.06%	
35) Dibromofluoromethane	7.57	113	216157	48.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.32%	
50) Toluene-d8	10.08	98	872735	50.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.18%	
62) 4-Bromofluorobenzene	12.40	95	288608	45.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.22%	
Target Compounds						
20) Methylene Chloride	4.53	84	83251	15.918	ug/l	99
43) Isopropyl Acetate	8.15	43	109951	9.003	ug/l #	90
95) Naphthalene	15.13	128	32905	4.479	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102119\
Data File : VN058886.D
Acq On : 21 Oct 2019 14:23
Operator : JC/SP
Sample : K5453-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RT1496

Quant Time: Oct 22 04:27:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

