

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102119\
 Data File : VN058887.D
 Acq On : 21 Oct 2019 14:45
 Operator : JC/SP
 Sample : K5447-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RT5124

Quant Time: Oct 22 04:30:13 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	394496	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	631095	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	580177	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	221812	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	287342	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
35) Dibromofluoromethane	7.57	113	204056	49.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.36%	
50) Toluene-d8	10.08	98	803243	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
62) 4-Bromofluorobenzene	12.40	95	260513	44.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.08%	
Target Compounds						
16) Acetone	3.80	43	33685	11.782	ug/l	98
18) Methyl Acetate	4.31	43	6948	1.069	ug/l #	61
20) Methylene Chloride	4.53	84	43242	8.858	ug/l	98
43) Isopropyl Acetate	8.15	43	105638	9.356	ug/l #	90
95) Naphthalene	15.13	128	68846	7.403	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102119\
Data File : VN058887.D
Acq On : 21 Oct 2019 14:45
Operator : JC/SP
Sample : K5447-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RT5124

Quant Time: Oct 22 04:30:13 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

