

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN102119\
 Data File : VN058889.D
 Acq On : 21 Oct 2019 15:29
 Operator : JC/SP
 Sample : K5147-07
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 OK-02-101819

Quant Time: Oct 22 04:34:57 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 19 00:16:20 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	392606	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	627975	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	591505	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	230057	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.01	65	287429	49.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.84%	
35) Dibromofluoromethane	7.57	113	201584	48.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.64%	
50) Toluene-d8	10.08	98	808849	50.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.94%	
62) 4-Bromofluorobenzene	12.40	95	272558	46.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.62%	
Target Compounds						
16) Acetone	3.80	43	14517	5.102	ug/l	94
18) Methyl Acetate	4.31	43	9963	1.540	ug/l #	70
20) Methylene Chloride	4.53	84	504602	103.859	ug/l	99
43) Isopropyl Acetate	8.15	43	106760	9.502	ug/l #	92
95) Naphthalene	15.13	128	81321	8.122	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN102119\
Data File : VN058889.D
Acq On : 21 Oct 2019 15:29
Operator : JC/SP
Sample : K5147-07
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
OK-02-101819

Quant Time: Oct 22 04:34:57 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N101819W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 19 00:16:20 2019
Response via : Initial Calibration

